



## Lab Report CS/ONH 21

# G4 ICARUS Series 2

- Fast and Reliable Carbon and Sulfur Determination in Limestone, Dolomite and Lime

### Introduction

Limestone and Dolomite are sedimentary rocks with a tremendous variety of applications in a wide variety of fields. Limestone is used as construction raw material, aggregate, fluxing agent, acid neutralizing agent, and an animal feed filler. When calcined, burnt lime (CaO) is created, that is more effective in acid-neutralization and a construction material. Dolomite is harder than limestone, making it a superior construction material. Quality- and Process-Control require the fast and accurate analysis of the elemental composition in these raw materials. This Lab Report shows the

simplicity, speed and reliability of carbon and sulfur determination by the G4 ICARUS Series 2 using a high-frequency (HF) induction furnace.

### Measuring Principle

The combustion takes place under a flow of oxygen within a sealed HF induction furnace. While the sample, with addition of a metallic accelerator, is heated by induction, it combusts at temperatures exceeding 2000 °C. Sulfur compounds are oxidized to SO<sub>2</sub> and carbon to CO<sub>2</sub>. Total carbon and sulfur are calculated relative to the amount of these combustion gases, which are quantified by selective detection systems.

## Combustion Without Compromises: G4 ICARUS!


With its powerful HF-furnace equipped with ZoneProtect™, its unique vacuum-free automatic cleaning system, HighSense™ detectors and electronic flow- and pressure-control, the G4 ICARUS Series 2 is a smart addition for every industrial user who depends on a reliable instrument even under harsh conditions. The G4 ICARUS Series 2 is capable to measure both carbon and sulfur in cements and related materials with high precision in typically below one minute.

### Sample Preparation

For optimal precision, the sample should be grinded or crushed to a uniform powder prior to analysis. If necessary, it should be dried for at least 1 h at 110 °C. No further sample preparation is necessary.

### Method Parameters

- Purge time: 5 s
- Start delay: 5 s
- Baseline check before analysis: 5 s
- Analysis time 1: 35 s (power Level: 4)
- Analysis time 2: 25 s (power level: 0)
- Baseline check after analysis: 5 s
- Crucible cool time: 10 s
  
- Sample mass: ~0.2 g (weighed to 0.1 mg)
- Accelerators:
  - Tungsten/Tin: ~1.5 g
  - High purity iron chips: ~0.7 g



### Calibration

The calibration of the analyzer is performed by means of reference materials with certified concentrations of carbon and sulfur. The analysis software supports single point calibration with one CRM and multi point calibration with multiple CRMs.

### Procedure

#### I. Determination of the blank value

Run a minimum of 3 analysis of the blank value by adding the described amount of accelerators into a preheated<sup>1)</sup> crucible and analyze.

#### II. Measuring reference materials

- Choose CRMs for calibration and define them in the analysis software with designation and the certified concentrations.
- Weigh in an appropriate amount of reference material into a preheated<sup>1)</sup> crucible, and transfer the exact mass into the analysis software. Cover the material with the described amount of accelerator and analyze.
- Repeat step 2 a minimum of three times for each reference material (or sample weight) used.

Calibrate the analyzer with the blank values recorded under I. and the results obtained with reference materials II. (for more details refer to the user manual).

#### III. Sample measurement

- Weigh ~0.2 g of sample into a preheated<sup>1)</sup> crucible and transfer the exact sample mass to the analysis software. Cover the sample with the described amount of accelerator and analyze.
- Repeat step 1 until an appropriate number of repetitions is obtained.

### Calibration Example

Typical concentration levels for total carbon and total sulfur in Limestone and Dolomite are in the range of Carbon: 10 - 15% and Sulfur: 0.02 - 0.4%, whereas the typical contents in burnt lime are Carbon: 0.1 - 2% and Sulfur: 0.01- 0.25%.

The typical concentration range is useful information to consider when selecting appropriate reference material for calibration. Although all detectors are linearized and an extrapolation of the calibration curve is possible, it is good laboratory practice that all samples fall within the calibrated range. For this Lab Report, the calibration and method is identical to that used for cement in Lab Report CS/ONH 020.

<sup>1)</sup> For optimal precision, ceramic crucibles shall be pre-heated in a muffle furnace at  $\geq 1250$  °C for a minimum of 15 min or  $\geq 1000$  °C for a minimum of 2 hours. To avoid contamination, crucibles must be handled with clean tongs and transferred to a desiccator for storage.

The resulting calibration for carbon and sulfur yield a  $R^2 = 0.9999$  which proves the excellent linearity, accuracy and reproducibility of the G4 ICARUS Series 2 detection system and the entire method.

Standard (lot#)	Material	Total carbon / %	Total sulfur / %
AR 4006 (411B)	ore	4.19	4.07
AR 4007 (1212E)	ore	7.27	3.26
AR 4012 (52199)	limestone	11.97	0.044
EZRM B 483-1	cast iron	2.46	0.103
Ag <sub>2</sub> SO <sub>4</sub> (p.a.)	pure substance	n.a.	10.12 – 10.28 <small>(98.5 - 100% purity)</small>

**Table 1:** Used reference materials

### Results

The reproducibility of the G4 ICARUS Series 2 and the method outlined is demonstrated by a series of 10 repetitive measurements of real production samples.

### Dolomite

Mass / g	Carbon / %	Sulfur / %
0.1754	13.10	0.023
0.1744	13.07	0.021
0.1754	13.10	0.023
0.1757	13.03	0.022
0.1754	13.07	0.023
0.1749	13.09	0.022
0.1766	13.02	0.021
0.1760	13.04	0.022
0.1743	13.06	0.023
0.1770	13.07	0.022
Mean	<b>13.06</b>	<b>0.022</b>
STD	0.03	<0.001

### Limestone

Mass / g	Carbon / %	Sulfur / %
0.2012	11.72	0.209
0.1957	11.73	0.210
0.1992	11.71	0.210
0.2061	11.72	0.206
0.1982	11.73	0.210
0.1889	11.71	0.209
0.1953	11.69	0.209
0.2004	11.73	0.207
0.2011	11.72	0.205
0.1978	11.72	0.210
Mean	<b>11.72</b>	<b>0.209</b>
STD	0.01	0.002



### Lime 1

Mass / g	Carbon / %	Sulfur / %
0.2001	0.426	0.244
0.2025	0.430	0.244
0.2012	0.434	0.248
0.2028	0.429	0.246
0.2018	0.429	0.245
0.2007	0.432	0.249
0.2019	0.428	0.248
0.2001	0.429	0.248
0.2013	0.429	0.250
0.2007	0.431	0.249
Mean	<b>0.430</b>	<b>0.247</b>
STD	0.002	0.002

### Lime 2

Mass / g	Carbon / %	Sulfur / %
0.2038	0.579	0.025
0.2027	0.579	0.025
0.2006	0.579	0.025
0.2003	0.578	0.026
0.2013	0.575	0.025
0.2018	0.576	0.023
0.2021	0.570	0.023
0.2016	0.581	0.023
0.2007	0.569	0.023
0.2050	0.570	0.024
Mean	<b>0.576</b>	<b>0.024</b>
STD	0.004	0.001

### Summary

The combination of HighSense™ detection systems with precise electronic flow control in the G4 ICARUS Series 2 delivers excellent and long-time stable analytical performance. The powerful HF-furnace equipped with the industry-leading ZoneProtect™ and its unique automatic cleaner ensures lowest cost of ownership, high availability with little maintenance. The sample preparation for this method is straightforward and does not require any expert knowledge. The analytical performance in combination with the speed of analysis and its ease of use makes the G4 ICARUS Series 2 the ideal choice for the quality control of limestone, dolomite, burnt lime and related products.



### ● Distribuce a servis v ČR a SR:

BAS Rudice spol. s r.o.

U Vodárny 1

678 01 Blansko

tel: +420 541 126 090

[www.bas.cz](http://www.bas.cz)